

https://twinkle.repo.nii.ac.jp

Effects of Low-Intensity Pulsed Ultrasound on Callus Formation: A Comparative Morphological Study

メタデータ	言語: eng
	出版者:
	公開日: 2018-01-30
	キーワード (Ja):
	キーワード (En):
	作成者: KONNO, Masayuki, ASANO, Hidetsugu, FUJII,
	Yusuke, KONISHI, Yoshiyuki, MURAGAKI, Yoshihiro
	メールアドレス:
	所属:
URL	http://hdl.handle.net/10470/00031743

Effects of Low-Intensity Pulsed Ultrasound on Callus Formation: A Comparative Morphological Study

Masayuki KONNO^{1,3}, Hidetsugu ASANO^{1,4}, Yusuke FUJII², Yoshiyuki KONISHI⁵ and Yoshihiro MURAGAKI⁵

¹Graduate School of Medicine, Tokyo Women's Medical University ²J. Morita-mfg. corp. ³Yours Dental Parkfield Clinic ⁴Pioneer Corporation ⁵Faculty of Advanced Techno-Surgery, Institute of Advanced Biomedical Engineering and Science, Tokyo Women's Medical University (Accepted June 5, 2017)

Ultrasound has a variety of applications as a noninvasive treatment. In particular, it is widely used for its accelerating effect on synostosis, through Low Intensity Pulsed Ultra-Sound (LIPUS). In this study, morphological evaluation was carried out of the effect of LIPUS on bone tissue, aimed at identifying the effective initiation timing for the treatment and the bone callus formation period.

A bone defect, 2 mm in both diameter and depth, was formed in the femur of 7-week-old male Sprague-Dawlay rats. 30 rats were assigned randomly into 5 groups: a group with no LIPUS stimulation (0 L); a group with LIPUS stimulation on day 1-3 (1-3 L); a group with LIPUS stimulation on day 1-14 (1-14 L); a group with LIPUS stimulation on day 4-14 (4-14 L); and a group with LIPUS stimulation for day 1-14 without a bone defect (Control). The inflammation period was considered to be 3 days. In order to evaluate the healing acceleration effect threedimensionally, comparison took place of hard tissue volume, as the number of high CT value voxels, and bone density as the average CT value. LIPUS stimulation during the inflammatory period was found to be of significant importance: the 1-14 L group's bone volume was significantly higher at day 7 and day 10 compared to the non-LIPUS stimulated groups in the inflammation period. This study suggests that it is desirable for LIPUS stimulation to be continued from the time of bone injury to the osteogenesis stage, and that LIPUS promotes callus formation, especially by stimulation in the inflammatory phase.

Key Words: LIPUS, ultrasound, cullus formation, CT, bone

Introduction

With the spread of minimally invasive treatments, focus has grown on the use of non-invasive ultrasound, which now has applications in a variety of treatments. Among these, there has been particular attention paid to the application of lowpower ultrasonic pulses (Low Intensity Pulsed Ultra-Sound: LIPUS). Reports based on animal experiments have suggested acceleration of both soft tissue¹⁾²⁾ and hard tissue healing³⁾⁻⁵⁾, with further ap-

⊠ : Masayuki KONNO Graduate School of Medicine, Tokyo Women's Medical University, 8–1 Kawada-cho, Shinjuku-ku, Tokyo, 162–8666 Japan

E-mail: yours.dental.konno@gmail.com

doi: 10.24488/jtwmu.87.4_108

Copyright © 2017 Society of Tokyo Women's Medical University

plication reported for use on the nervous system⁶. LIPUS treatment has even been reported for use within the osteoporosis model⁷⁾⁸⁾. Analysis at a genetic level has also identified physical stimulation through LIPUS as promoting cell differentiation of osteoblasts and chondrocytes ^{9)~11)}. Furthermore, there have also been reports of increased mechanical strength¹²⁾ and bone mineral content¹³⁾ resulting from LIPUS stimulation. The results of LIPUS stimulation on bone union are clear and the application of LIPUS to aid post-fracture recovery has now become widespread. In addition to its utilization for post-fracture treatment, LIPUS application on implants has been reported to shorten the healing period^{14)~18)}. Although there is research regarding the optimal working conditions for LIPUS frequency¹⁹, currently the only safe and effective conditions for fracture healing are a frequency of 1.5 MHz, a repetition frequency of 1.0 kHz, an ultrasound intensity of 30 mW/cm², a pulse width of 200 μ s, and stimulation time of 20 minutes/day²⁰⁾.

LIPUS has been approved as a medical tool since the 1990s and is used in a variety of applications, such as bone fracture treatment. U.S Clinical trials have acknowledged the effects of LIPUS stimulation in the acceleration of bone union²¹⁾²²⁾. While there exist studies looking at the optimal period for LIPUS stimulation²³⁾, review of these studies related to bone union following LIPUS stimulation shows that bone evaluation was predominantly done through assessment of mechanical strength²⁴⁾, using two-dimensional imaging²⁵⁾. Within these studies there was also a lack of three-dimensional (3D) morphological evaluation of bone tissue. The fracture healing phase is made up of the inflammatory phase, the cell growth phase, the callus formation phase and the remodeling phase, with the callus formation phase reached between week 1 and week $4^{2^{6})^{27}}$. However, the difference between the acceleration effect of LIPUS stimulation at each of these phases is unclear and there are also no reports on the duration or period of LIPUS stimulation necessary for effective bone union. There is, therefore, a possibility that more effective post-bone fracture healing acceleration may be achieved through LI- PUS stimulation treatment if it is applied over an optimal period and duration. In this study, 3D morphological assessment is used on LIPUS stimulation on rats with the bone fracture model. The objective was to clearly show the optimal stimulation period and duration for post-fracture LIPUS treatment.

Materials and Methods

The experiment was done using 7-week-old, male Sprague-Dawlay rats (SD rats, Charles River). The entire bone fracture healing phase is made up of the inflammatory phase, the cell growth phase, the callus formation phase and the remodeling phase²⁶⁾. However, in the pilot experiment the period necessary for the femur of the SD rats to recover enough strength was approximately two weeks. Therefore, based on this, the experiment duration was set to run over 15 days to reach the callus formation phase.

Preparation of the fracture model rats aimed for stable reproduction of the fracture site and was conducted using a drill (GC Corp., Depth Drill 6B341) to create a cavity diameter of 2 mm and a depth of 2 mm at the injury site of the femur. The drill was cooled between procedures in order to prevent burning and was thoroughly flushed with saline to ensure no fragments of bone remained.

In order to investigate the relationship between LIPUS stimulation period and post-bone damage bone union, with reference to Nakajima et al²⁸, the inflammatory phase was set at three days and the experiment conducted with the subjects split between the five groups below.

1. Bone damage procedure applied, LIPUS stimulation not carried out (0 L group)

2. Bone damage procedure applied, LIPUS stimulation carried out from day 1-3 (1-3 L group)

3. Bone damage procedure applied, LIPUS stimulation carried out from day 1-14 (1-14 L group)

4. Bone damage procedure applied, LIPUS stimulation carried out from day 4-14 (4-14 L group)

5. Control group. Bone damage procedure not applied, LIPUS stimulation carried out from day 1-14 (Control group)

Each of the five groups contained six rats, making a total of 30 rats. (The day of the bone damage




Fig. 1 LIPUS stimulation conditions (A) is the LIPUS stimulation head unit.

procedure was counted as day 0).

LIPUS stimulation of the rats was carried out from day lor day 4 and was conducted under anesthesia. During treatment the SAFHS LIPUS stimulation equipment was used (Teijin Pharma Ltd., SAFHS 4,000 J). Stimulation was carried out with a frequency of 1.5 MHz, a repetition frequency of 1.0 kHz, an ultrasound intensity of 30 mW/cm², a pulse width of 200 µs and a stimulation time of 20 minutes per day⁹. As shown in Fig. 1, during treatment the ultrasonic wave stimulation unit was affixed to the femur of the anesthetized rat using surgical tape and stimulation was then carried out. Shaving of the target site was conducted before the bone damage procedure. Shaving was repeated regularly to enable accurate repeat affixing of the stimulation head. X-ray micro-CT imaging of the rat femurs was performed for all groups on days 1, 3, 7, 10 and 15. The CT imaging was performed using the 3D micro-X-ray system R_mCT2 (manufactured by the Rigaku Corporation). The resolution of the CT device was 59-µm and the scan was carried out with an FOV diameter of 30 mm, a height of 30 mm, and an exposure time of two minutes.

The experiment received the permission of the animal experimentation ethics committee of Tokyo Women's Medical University (15-23) and was carried out in compliance with its standards.

Evaluation was carried out based on the CT values. A CT value of 200HU was set as the threshold for soft tissue: values below this threshold were excluded from the evaluation. Bone volume was calculated as the number of voxels. In addition, bone density was calculated as the sum of the CT values divided by the number of voxels, that is, the average CT value. It should be noted that due to the properties of ultrasound, LIPUS stimulation can be reflected and refracted depending on the tissue boundary surface. This is particularly the case for hard tissue boundaries, meaning the actual site stimulated by ultrasound is limited to the part facing the stimulation unit²⁹⁾³⁰⁾. In consideration of this, in the analysis of this study, the femur was divided into halves on the LIPUS stimulation surface and non-stimulation surface, and the volume of each bone and the density of bone were measured. Specifically, as shown in A in Fig. 2, the central axis of the hole was set as the x axis, the central axis of the femur was set as the y axis, and the z axis formed the right hand coordinate system defined by the x axis and the y axis. At this time, it was cut out into a cylindrical shape with a diameter of 9.4 mm and a height of 2.4 mm with the y axis as the central axis and the xz plane as the object plane, and it was divided into two on the yz plane, LIPUS stimulation side (Fig. 2B) and non-stimulation side (Fig. 2C).

Analysis of variance (ANOVA) and multiple comparison tests (Tukey-Kramer method) were applied on the measurement results from each group: results from the LIPUS stimulation side and nonstimulation side together and results solely from the LIPUS stimulation side (10 groups in total).

Results

Fig. 3(a) shows the bone volume results for the LIPUS stimulation side.

Over the LIPUS stimulation period no clear change in bone volume of the control group was ob-




Fig. 2 An image obtained by Micro CT A is a 3D reconstructed bone image. B shows the analyzed area including the bone defect. C is the other side of B.

served. Within the other groups, an increase of bone volume for bone union was confirmed regardless of the presence or absence of LIPUS stimulation or the period of the stimulation.

Comparing the results at day 1, the bone volume of the control group was significantly higher than that of the bone-damaged groups (p < 0.05), but there was no clear difference among the bonedamaged groups. The results for day 3 were the same as for day 1; however the differences between the bone-damaged groups and the control group had become smaller. By day 7, the bone volume of the bone-damaged groups had become higher than that of the control group. A trend was seen in the bone-damaged groups that underwent LIPUS stimulation during the inflammatory period (1-3 L group, 1-14 L group), with higher bone volume compared to the other bone-damaged groups (0 L group, 4-14 L group). In particular, the bone volume of the 1-14 L group (44,933 ± 2,958 voxels) was significantly higher (p < 0.05) in comparison to the 0 L group (36,288 ± 4,490 voxels) and 4-14 L group $(37,035 \pm 3,962 \text{ voxels})$. The same trend seen at day 7 was also apparent at day 10: the bone volume of the 1-14 L group (48,613 \pm 3,738 voxels) was significantly higher (p < 0.05) than the 0 L group (40,461 \pm 5,157 voxels) and the 4-14 L group (41,599 \pm 2,354 voxels). The differences between the bone-damaged groups and the control group had also increased. By day 15, the difference in bone volume among the bone-damaged groups and the control group was significantly smaller (p < 0.05); however there was no significant difference between any combination of the bone-damaged groups.

Fig. 3(b) shows the transition result of the bone density for the LIPUS stimulation side.

At day 1 no differences in bone density were observed between each pair of groups. By day 3 differences in bone density between the control group and the bone-damaged groups had begun to appear, the bone-damaged groups having lower values than the control group and being relatively soft. At day 7, the average bone density of the control group was 505 \pm 8, significantly different (p < 0.05) to those of the bone-damaged groups: the 0 L group (403 \pm 29 HU), the 1-3 L group (397 \pm 46 HU), the 1-14 L group (423 \pm 20 HU), and the 4-14 L group (421 \pm 18 HU). The results at day 10 were the same as for day 7 and a significant difference was again observed (p < 0.05): the average bone density of the 32




Fig. 3 Time series graph showing bone volume (a) and bone CT value (b) (LIPUS stimulated side).

control group was 515 ± 22 , while the average density of the remaining bone-damaged groups showed no substantial change. By day 15, difference in average bone density among the groups was almost no longer apparent.

It should be noted that no clear change was observed in bone volume or bone density of the non-LIPUS stimulated side.

Fig. 4 shows a typical example of the micro-CT images from the 0 L and 1-14 L groups. A and B show the micro-CT images of the 0 L group on day 1 and day 7 respectively. D and E are the CT images of the 1-14 L group on day 1 and day 7. Images C and F are 3D diagrams of the CT images from the

0 L group and the 1-14 L group at day 7. Comparing B and E, it can be seen that by day 7 both cancellous bone impermeability and hard bone volume of the 1-14 L group had increased. It is also apparent from the 3D images that compared to C (0 L group) bone formation can be clearly confirmed at the bone injury site in F (1-14 L group).

Discussion

The experimental results show that at day 1 and day 3, although bone volume for the non-bonedamaged control group was significantly higher, no difference was observed among the bone-damaged groups. This difference in comparison to the control group is attributed to being due to the volume of




Fig. 4 Comparison between an LIPUS non-stimulated rat and a stimulated rat A and B are CT images from an LIPUS non-stimulated rat (0 L). A is a cross-sectional image on day 1. B is a cross-sectional image on day 7. C is a 3D reconstructed image of B. D-F show the LIPUS stimulated rat data (1-14 L).

bone removed as part of the bone damage treatment. By day 7 and day 10 the bone volume of the 1-14 L group, which underwent continued LIPUS stimulation, had greatly increased. In particular, a significant difference in bone volume was seen when compared to groups that had not undergone LIPUS stimulation during the inflammatory phase (0 L group, 4-14 L group). Based on this difference, LIPUS stimulation during the inflammatory phase is thought to be an important factor in accelerating bone recovery. The 1-3 L group, which had LIPUS stimulation only during the inflammatory phase, showed a tendency towards higher bone volume compared to the 0 L group; however, a significant difference was not seen. In regards to bone volume, even if LIPUS stimulation is only applied during the inflammatory phase, there may be a small healing acceleration effect. At day 15, significant differences in bone volume was observed between the bonedamaged groups and the control group: however no difference was found among the bone-damaged groups themselves. From these results it can be considered that application of LIPUS immediately after bone damage accelerates bone volume increase during the early period; however, after callus formation bone volume becomes roughly equal regardless of the presence, duration or period of LI-PUS stimulation.

On the other hand, the bone density results show that at day 7 and day 10, the bone-damaged groups had a lower average bone density compared to the control group, suggesting the formation of immature bone as part of the bone healing process. However, clear differences among the bone-damaged groups themselves were not observed. At day 15 there was no longer a difference in average bone density among the groups, suggesting LIPUS stimulation does not affect bone density after healing.

The results above suggest that after bone injury, LIPUS stimulation in the inflammatory phase and the cell growth phase accelerate the formation of immature bone, speeding up bone recovery in the early period. Furthermore, LIPUS stimulation does not affect bone volume or bone density at the injury site after healing and has little effect on normal bone. Therefore LIPUS should be used only during the healing process. Initiating LIPUS stimulation in the early period following bone injury, during the inflammatory phase and cell growth phase before the callus formation phase, is expected to accelerate bone union effectively. Also, while Azuma et al²³⁾ reported an increase in mechanical strength resulting from LIPUS, the results from this experiment suggest an increase in mechanical strength could occur due to an increase in bone volume, without a change in bone density.

It also has been reported by Jingusi et al¹⁵⁾ and Iwabuchi et al¹⁶⁾ that as LIPUS stimulation is dampened by soft tissue the position of the stimulation head directly affects the therapeutic effect. In this experiment using rats, the comparison results from the LIPUS stimulation side and non-stimulation side suggest that healing acceleration is limited to the LIPUS stimulation side. Sasaki et al¹⁹⁾ reported that the LIPUS stimulation accelerates capillary vessel formation in the LIPUS stimulation side. Acceleration of capillary vessel formation is limited to the LI-PUS stimulation side, suggesting that there is a relationship with our study. In terms of human application, for example in the case of a complete fracture, there is a possibility that further healing acceleration effects may be obtained through LIPUS stimulation of the entire fracture site.

LIPUS treatment has been reported as affecting osteoblasts, osteoclasts, cartilage cells and mesenchymal cells³¹⁾, but the specific mechanisms, stimulation times and duration involved remain unclear. This experiment suggests the optimal stimulation time as being the period before callus formation.

In the future, similar verification of reports that LIPUS stimulation is different on mature rats²⁵⁾ is necessary to seek a more effective method of treatment. In addition to this, 3D morphological evaluation of bone tissue and implant treatment after tooth extraction is required to investigate if a similar bone regeneration effect can also be expected.

Conclusion

This study evaluated the callus formation acceleration effect of LIPUS stimulation, from the point of bone damage to callus formation, using CT imaging for 3D morphological evaluation, in order to verify the effective LIPUS stimulation period and duration. The group that underwent continuous stimulation directly following bone damage showed an increased bone volume in the early period compared to the other groups. This suggests that continuous LIPUS stimulation after bone injury occurs is effective in the acceleration of callus formation.

Acknowledgments

We would like to thank Yoshiharu Kato, Professor of Tokyo Women's Medical University, for his valuable advice, including regarding the experimental methods, as well as Masanori Maeda and Manabu Tamura, Assistant Professor of Tokyo Women's Medical University Institute of Advanced Biomedical Engineering and Science for their support in creating this paper.

During the LIPUS stimulation experiment, free loan of the SAFES 4000J machine was provided by the Teijin Pharma Corporation.

References

- Maeda T, Masaki C, Kanao M et al: Low-intensity pulsed ultrasound enhances palatal mucosa wound healing in rats. J Prosthodont Res 57: 93–98, 2013
- Nishimura S, Tatsumi J, Hayashi K et al: Effect of low-intensity pulsed ultrasound on wound healing after periodontal surgery. The Journal of Meikai Dental Medicine 42: 98–109, 2013 (in Japanese)
- Duarte LR: The stimulation of bone growth by ultrasound. Arch Orthop Trauma Surg 101: 153–159, 1983
- Pilla AA, Mont MA, Nasser PR et al: Noninvasive low-intensity pulsed ultrasound accelerates bone healing in the rabbit. J Orthop Trauma 4: 246–253, 1990
- 5) Wang F, Li Y, Yang Z et al: Effect of low-intensity pulsed ultrasound on a rat model of dentin-dental pulp injury and repair. Ultrasound Med Biol 43: 163–175, 2017
- 6) Sato M, Motoyoshi M, Shinoda M et al: Lowintensity pulsed ultrasound accelerates nerve regeneration following inferior alveolar nerve transection in rats. Eur J Oral Sci 124: 246–250, 2016
- Wu S, Kawahara Y, Manabe T et al: Lowintensity pulsed ultrasound accelerates osteoblast differentiation and promotes bone formation in an

osteoporosis rat model. Pathobiology **76**: 99–107, 2009

- Cheung WH, Chin WC, Qin L et al: Low intensity pulsed ultrasound enhances fracture healing in both ovariectomy-induced osteoporotic and agematched normal bones. J Orthop Res 30: 129–136, 2012
- 9) Mukai S, Ito H, Nakagawa Y et al: Transforming growth factor-beta1 mediates the effects of lowintensity pulsed ultrasound in chondrocytes. Ultrasound Med Biol 31: 1713–1721, 2005
- 10) Kidokoro T, Takeuchi K, Murakami H et al: Effects of low-intensity pulsed ultrasound on osteoblastic cells derived from rat bone marrow. J Jpn Soc Oral Implant 20: 450–458, 2007 (in Japanese)
- 11) Suzuki A, Takayama T, Suzuki N et al: Daily lowintensity pulsed ultrasound-mediated osteogenic differentiation in rat osteoblasts. Acta Biochim et Biophys Sin 41: 108–115, 2009
- 12) Jingusi S: Effects of low-power ultrasonic pulse irradiation on femur fracture healing in rats. J Jpn Soc Fracture Repair 21: 655–658, 1999 (in Japanese)
- 13) Mayr E, Laule A, Suger G et al: Radiographic results of callus distraction aided by pulsed lowintensity ultrasound. J Orthop Trauma 15: 407–414, 2001
- 14) Ganzorig K, Kuroda S, Maeda Y et al: Lowintensity pulsed ultrasound enhances bone formation around miniscrew implants. Arch Oral Biol 60: 902–910, 2015
- 15) Liu Q, Liu X, Liu B et al: The effect of lowintensity pulsed ultrasound on the osseointegration of titanium dental implants. Br J Oral Maxillofac Surg 50: 244–250, 2012
- 16) Nakanishi Y, Wang P-L, Ochi M et al: Lowintensity pulsed ultrasound stimulation significantly enhances the promotion of bone formation around dental implants. J Hard Tissue Biol 20: 139– 145, 2011
- 17) Zhou H, Hou Y, Zhu Z et al: Effect of low-intensity pulsed ultrasound on implant osseointegration in ovariectomized rats. J Ultrasound Med 35: 747–754, 2016
- 18) Naka T, Yokose S: Effect of low-intensity pulsed ultrasound on osseointegration in a rat model. J Jpn Soc Oral Implant 25: 31–39, 2012 (in Japanese)
- 19) Sasaki H, Monden K, Yoshinari M et al: Comparison of angiogenesis in bone defect healing process

due to the difference in the frequency of lowintensity pulsed ultrasound (LIPUS). J Hard Tissue Biol **25**: 157–164, 2016

- 20) **Jinguji S, Matsushita** T: Basic and Clinical Application of Low-intensity Pulse Ultrasound Treatment for Fractures, medical review, Osaka (2008)
- 21) Heckman JD, Ryaby JP, McCabe J et al: Acceleration of tibial fracture-healing by non-invasive, low-intensity pulsed ultrasound. J Bone Joint Surg Am 76: 26–34, 1994
- 22) Kristiansen TK, Ryaby JP, McCabe J et al: Accelerated healing of distal radial fractures with the use of specific, low-intensity ultrasound. A multicenter, prospective, randomized, double-blind, placebo-controlled study. J Bone Joint Surg Am 79: 961–973, 1997
- 23) Azuma Y, Ito M, Harada Y et al: Low-intensity pulsed ultrasound accelerates rat femoral fracture healing by acting on the various cellular reactions in the fracture callus. J Bone Miner Res 16: 671–680, 2001
- 24) Warden SJ, Fuchs RK, Kessler CK et al: Ultrasound produced by a conventional therapeutic ultrasound unit accelerates fracture repair. Phys Ther 86: 1118–1127, 2006
- 25) Yoshida A, Sasaki H, Furuya Y et al: Effect of low-intensity pulsed ultrasound on bone-healing process in murine low-turnover osteoporosis model. J Hard Tissue Biol 22: 301–309, 2013
- 26) Bone Biology, (2nd ed). (Suda R, Ozawa H, Takahashi H eds), Ishiyaku pub, Tokyo (2016)
- 27) Matsumoto Dental University Graduate School of hard tissue research group: Hard Tissue Research Handbook, Matsumoto Dental University Press, Nagano (2008)
- 28) Nakajima A, Yamazaki M, Takahashi K: Recent progress in fractured healing research through molecular and cellular biology. Chiba Medical Journal 86: 83–91, 2010 (in Japanese)
- 29) Jinguji S: Basic and clinical fracture treatment using low-power ultrasonic pulse. Orthopedics Orthopedic Surgery 51: 1471–1477, 2000 (in Japanese)
- 30) Iwabuchi T: Medical ultrasound analysis of ultrasonic wave transmission of the femoral intramedullary fixed model. Ultrasound Techno 18: 32–36, 2006
- Claes L, Willie B: The enhancement of bone regeneration by ultrasound. Prog Biophys Mol Biol 93: 384–398, 2007

低出力超音波パルス照射における仮骨形成への影響の形態的比較に関する研究

¹東京女子医科大学大学院医学系研究科 ²株式会社モリタ ³ユアーズ歯科パークフィールドクリニック ⁴パイオニア株式会社

*東京女子医科大学先端生命医科学研究所先端工学外科学分野(指導:村垣善浩教授)

コンノ マサユキ アサノ ビデツグ ユウスケ コニシ ヨシユキ ムラガキ ヨシヒロ 今野 雅之¹³・浅野 秀胤¹⁴・藤井 優輔²・小西 良幸⁵・村垣 善浩⁵

超音波はその非侵襲性に注目が集まり,様々な治療に用いられるようになってきた.中でも低出力超音波パルス (Low Intensity Pulsed Ultra-Sound:LIPUS)による骨癒合促進効果が広く利用されている.しかしながらLIPUS 照射期間と時期の仮骨形成に関する研究では、骨の機械的強度やX線画像による評価が主であり、骨組織の3 次元的な形態評価は少ない。本研究は LIPUS 照射による仮骨形成を3次元的に形態評価し、骨癒合に最適な照射 時期および期間を明らかにする.7週齢の雄SDラットを使用し実験を行った.大腿骨にドリルを用いて直径2 mm, 深さ2mmの骨損傷を形成した.炎症期が3日であることを考慮し、5群に分け実験を行った.骨損傷処置 を行い LIPUS 照射を行わなかった群(0L群). 骨損傷処置を行い Day 1~3 に LIPUS 照射した群(1-3 L 群), 骨 損傷処置を行い Day 1~14 に LIPUS 照射した群 (1-14 L 群), 骨損傷処置を行い Day 4~14 に LIPUS 照射した群 (414 L 群), コントロールとして骨損傷処置を行わずに Day 1~14 に LIPUS 照射した群 (Cont 群), 各群 6 匹の 計 30 匹で,実験日数は 15 日とした. LIPUS 照射は周波数 1.5 MHz,繰り返し周波数 1.0 kHz,超音波強度 30 mW/ cm², パルス幅 200 μs, 照射時間は1日あたり 20 分間とし, 全群に対して Day 1, 3, 7, 10, 15 にラット大腿骨の X 線マイクロ CT 撮影を行った. 骨の癒合促進効果を3次元的に評価するために, CT 値を元に軟組織を除く硬組織 のボクセルデータの数を体積とみなした骨体積と平均 CT 値による骨の密度を比較した. Day 7, Day 10 では LI-PUS を当て続けた1-14L 群の骨体積が大きく、特に炎症期における LIPUS 照射のない群との比較では有意に骨 体積が大きいことから、炎症期の LIPUS 照射が特に重要であることがわかった.本研究より LIPUS 照射は、骨損 傷直後から仮骨形成期までの間に照射し続けることが望ましく、LIPUS は特に炎症期に照射することで仮骨形成 を促進させることが示唆された。