

https://twinkle.repo.nii.ac.jp

Adenosine Triphosphate Bioluminescence Test of the Nasal Spray Nozzles Attached to an Ear-nose-throat Treatment Unit

| メタデータ | 言語: eng                                   |
|-------|---|
|       | 出版者:                                      |
|       | 公開日: 2019-03-15                           |
|       | キーワード (Ja):                               |
|       | キーワード (En):                               |
|       | 作成者: YAMAMURA, Yukie, EGAWA, Tomoko, SEO, |
|       | Yukako, NONAKA, Manabu                    |
|       | メールアドレス:                                  |
|       | 所属:                                       |
| URL   | http://hdl.handle.net/10470/00032117      |

1

# Adenosine Triphosphate Bioluminescence Test of the Nasal Spray Nozzles Attached to an Ear-nose-throat Treatment Unit

# Yukie YAMAMURA<sup>1</sup>, Tomoko EGAWA<sup>2</sup>, Yukako SEO<sup>1</sup>, and Manabu NONAKA<sup>1</sup>

<sup>1</sup>Department of Otolaryngology, Tokyo Women's Medical University, Tokyo, Japan <sup>2</sup>3rd Grade Medical Student, School of Medicine, Tokyo Women's Medical University, Tokyo, Japan (Accepted October 31, 2017) (Advance Publication by J-STAGE January 20, 2018)

To report a proper protocol for cleaning and disinfecting the nozzles attached to ear-nose-throat (ENT) treatment unit, contamination of nasal spray nozzle, nasal speculum, and suction tube before and after use was assessed by adenosine triphosphate (ATP) bioluminescence test. The measured ATP was expressed as relative light units (RLU).

There were no statistical differences in RLU scores among the nasal spray nozzles, nasal speculums, and suction tubes before patient use. Nozzle RLU scores increased after patient use. The RLU scores (mean  $\pm$  SD) of the nasal spray nozzles before use was 12  $\pm$  7.1, increasing to 30  $\pm$  28.7 after patient use. The RLU scores of the nasal speculums increased to 9,722  $\pm$  12,398.9 after patient use. The RLU scores of the suction tubes increased to 88,366  $\pm$  106,839.3 after patient use. Increased RLU scores of spray nozzles were statistically lower than those of nasal speculums and suction tubes.

In conclusion, nasal spray nozzles should be wiped with a mid-level disinfectant after use. Alternatively, the nozzle tips should be changed between patients. These recommendations should be considered when developing protocols for cleaning and disinfection of nasal spray nozzles used in conjunction with ENT treatment units.

Key Words: ENT treatment unit, spray nozzle, contamination, ATP bioluminescence test

# Introduction

Nasal spray nozzles attached to ear-nose-throat (ENT) treatment units are not routinely changed between patient examinations because the nozzles do not directly contact patients' skin or mucosa. However, there is a risk of cross-contamination by accidental touching of the spray tip to the mucosa, or by contact with nasal discharge, droplets, or aerosolized material because the spray is used near the nostrils (**Fig. 1**). Past studies have reported the detection of coagulase-negative staphylococcus

(CNS), methicillin-sensitive *Staphylococcus aureus* (MSSA), and sometimes methicillin-resistant *S. aureus* (MRSA) on the surface of nasal spray nozzles following their use<sup>1)2)</sup>.

Microbiological culturing is a traditional method for hygiene monitoring; however, these methods have some limitations. They require several days to get results because a colony growth incubation period is needed. Moreover, these methods cannot detect nonbacterial contaminants such as blood and mucous.

Recently, adenosine triphosphate (ATP)-based micro-

Corresponding Author: Yukie Yamamura. Department of Otolaryngology, Tokyo Women's Medical University, 8-1 Kawada-cho, Shinjuku-ku, Tokyo 162-8666, Japan

E-mail: yamamura.yukie@twmu.ac.jp

doi: 10.24488/twmuj.2017002

uol. 10.24488/twilluj.2017002

Copyright © 2018 Society of Tokyo Women's Medical University



Fig. 1 Using nasal spray near the nostrils of patient

biologic monitoring methods were developed for hygiene monitoring<sup>33</sup>. These methods measure ATP, the principal carrier of energy in all living organisms including microorganisms. ATP testing can be completed within 10 seconds, and is considerably easier and produces more rapid results compared to traditional swabbing culture methods. We assess the presence of contaminants on the nasal spray nozzle surfaces attached to ENT treatment units via ATP test to report a proper protocol for cleaning and disinfecting the nozzles.

# **Materials and Methods**

The study was conducted in December 2016. We examined six spray nozzles attached to three ENT treatment units located in the consultation room of an ENT practice. Each unit was used to administer two sprays containing 2% lidocaine and 0.02% adrenaline (**Fig. 2**). As per the clinic's standard operating procedures, the spray nozzles were cleaned using hot (90°C) water once per week prior to being reattached to the ENT treatment units.

# 1. ATP bioluminescence test

Prior to testing, the spray nozzles were wiped with cotton soaked in isopropanol. The 3 M<sup>TM</sup> Clean-Trace<sup>TM</sup> Surface ATP test system was used, consisting of a test swab, cuvette, and 3M<sup>TM</sup> Clean-Trace<sup>TM</sup> NG Luminometer. The ATP testing protocol was performed according



Fig. 2 Nasal spray nozzles attached to ENT treatment unit (arrows: tip)

to the manufacturer's instructions<sup>4)</sup>. The tips of the nozzles were rubbed with a test swab. The swabs were then inserted into a cuvette containing a luciferin, luciferase and Mg<sup>2+</sup>, causing the following enzymatic reaction:

ATP + luciferin/luciferase→adenosine monophosphate (AMP) and phosphoric residues+ light (wave of 562 nm)

The intensity of this light is proportional to the amount of ATP on the object surface, which corresponds to the amount of contamination by organic residues and microorganisms. Measurement of light intensity requires a Luminometer. This device contains a measuring chamber isolated from external light sources, and a detector that processes the optic signal to the electrical signal, which is expressed in relative light units (RLU)<sup>4)</sup>. Baseline RLU scores were recorded as "before use" scores.

The spray nozzles were then used on three patients, and ATP testing was repeated, with the resultant RLU scores recorded as "after use" scores. This cycle of nozzle cleaning, baseline ATP testing, nozzle use with three patients, and repeat ATP testing was performed 4-6 times per each spray nozzle, for a total of 28 times. For comparison, we also obtained RLU scores from the outer surfaces of nasal speculums, which were in direct contact with the nasal cavity, and suction tubes, which were inserted into the nasal cavity.

All statistical analyses were performed using Graph-Pad Prism version 6.0 for Windows. Paired t-tests were used to compare spray nozzle RLU scores obtained before patient use (after cleaning) and after patient use. A one way analysis of variance and Tukey-Kramer HSD



**Fig. 3** Comparison of RLU scores between nasal spray nozzles, nasal speculums and suction tubes before use

There were no statistical differences in RLU scores among the nasal spray nozzles, nasal speculums, and suction tubes before patient use.



**Fig. 4** RLU scores of nasal spray nozzles before and after use Nozzle RLU scores increased after patient use. The RLU scores (mean  $\pm$  SD) of the nasal spray nozzles before use was  $12 \pm 7.1$ (highest RLU = 25), increasing to  $30 \pm 28.7$  (highest RLU = 164) after patient use.

tests were used to compare the logarithm of RLU scores among the spray nozzles, nasal speculums, and suction tubes. Statistical significance was set at a p-value of < 0.05.

### 2. Culturing methods

We obtained 10 samples from the surfaces of nasal



Fig. 5 Comparison of logarithm of RLU scores between nasal spray nozzles, nasal speculums and suction tubes after use The RLU scores (mean  $\pm$  SD) after patient use is: nasal spray nozzles,  $30 \pm 28.7$ ;

nasal speculums,  $9,722 \pm 12,398.9$ ;

suction tubes,  $88,366 \pm 106,839.3$ .

Increases in RLU scores of spray nozzles were statistically smaller than the increases observed with the nasal speculums and suction tubes.

spray nozzles used with the three patients for culturing using the Transwab<sup>™</sup> ENT Amies charcoal MWE. Routine laboratory procedures were used to identify bacterial species on the surface of the nozzles. All growths were considered microbiologically positive, regardless of species or number of colonies forming the unit.

# Results

### 1. ATP bioluminescence test

There were no statistical differences in RLU scores among the nasal spray nozzles, nasal speculums, and suction tubes before patient use (**Fig. 3**). Nozzle RLU scores increased after patient use. The average RLU score (mean  $\pm$  SD) of the nasal spray nozzles before use was  $12 \pm 7.1$ , increasing to  $30 \pm 28.7$  after patient use (**Fig. 4**). The average RLU score of the nasal speculums increased from  $11 \pm 2.8$  to  $9,722 \pm 12,398.9$  after patient use. The average RLU score of the suction tubes increased from  $13 \pm 5.7$  to  $88,366 \pm 106,839.3$  after patient use. Increases in RLU scores of spray nozzles were statistically smaller than the increases observed with the nasal speculums and suction tubes (**Fig. 5**).

#### 2. Culturing methods

Microbiological cultures of all 10 samples were negative for bacterial growth.

# Discussion

According to Spaulding<sup>5)</sup>, medical devices are classified into three categories based on the risk of infection associated with use: critical, semi-critical, and noncritical. Critical devices contact sterile tissue or vasculature and should therefore be sterilized. Semi-critical devices contact intact mucous membranes, but do not penetrate sterile tissue, and therefore should be cleaned with a highlevel disinfectant or undergo hot water disinfection. Noncritical devices touch only intact skin or do not touch the patient at all. These devices should be cleaned using a low-level disinfectant.

Nasal spray nozzles are classified as noncritical devices because, in typical practice, they do not touch the skin or mucosa. However, our study demonstrated increased RLU scores following patient use, suggesting contamination originating from the aerosol or accidental touching with patient's mucus, or nasal discharge. RLU scores obtained from the spray nozzles were low compared with those obtained from the nasal speculums and suction tubes, which directly contacted patient skin or mucosa.

In the present study, all samples obtained from the spray nozzles were negative for conventional (swabbased) tests of bacterial growth. However, previous studies reported normal bacterial flora obtained from spray nozzle tips, including CNS, MSSA, and MRSA. Organisms that routinely cause respiratory tract infections such as *Haemophilus influenzae*, *Streptococcus pneumoniae*, or *Pseudomonas aeruginosa* were not detected<sup>1)2/6)</sup>.

Nasal spray nozzles should be wiped using a mid-level disinfectant such as ethanol or isopropanol following patient use. Nasal sprays that use disposable tips are also useful.

ATP tests for hygiene monitoring are well-established in food production facilities, with additional applications for monitoring clinical devices and environments<sup>30</sup>. A study assessing hospital kitchen surfaces showed a statistically significant relationship between ATP testing and traditional cultures obtained via microbial swabbing<sup>7</sup>. ATP testing can be completed easily and rapidly compared with traditional swabbing culture methods. However, ATP tests are not a substitute for culturing methods because the ATP test is unable to determine the identities of specific bacterial strains. Furthermore, ATP testing sensitivity varies among substrates. Turner et al.<sup>8</sup> reported that pure bacteria are weakly detected with an limit of detection of 10<sup>4</sup> for representative gram-negative and 10<sup>2</sup> for representative gram-positive bacteria. ATP testing is unable to detect the gram-negative bacteria because of incomplete cell lysis.

ATP tests can determine the general form of surface contamination, including cultivable and non-cultivable microbial and organic contamination. In the present study, ATP testing detected mild contamination on the surface of the nasal spray nozzles. These results would not have been obtained using traditional swabbing cultural methods.

# Conclusion

We measured contamination of the surfaces of nasal spray nozzles using ATP bioluminescence testing. After use with three patients, the nasal spray nozzles showed increased RLU scores, indicative of contamination, although RLU scores obtained from nasal spray nozzles were statistically lower than those obtained from nasal specula and suction tubes, both of which directly contact the nasal cavity or nasal discharge. Nasal spray nozzles should be wiped with a mid-level disinfectant after use. Alternatively, the nozzle tips should be changed between patients. These recommendations should be considered when developing protocols for cleaning and disinfection of nasal spray nozzles used in conjunction with ENT treatment units.

# Acknowledgements

This study was performed as a part of Research Project for 3rd grade (Segment 6) Medical Students at Tokyo Women's Medical University. The authors would like to thank Prof. Hiroki Fujieda and the research project committee members.

Conflicts of Interest: The authors have no conflicts of in-

terest, and no financial disclosures to report.

#### References

- Mori S: Quantitative evaluation of bacterial contamination of the otorhinolaryngologic medical instruments without thorough disinfection. J Jpn Soc Infect Aerosol Otorhinolaryngol 4: 14–19, 2016 (in Japanese)
- Nakajima N, Abe M, Miwa M et al: Bacterial investigation of nasal spray nozzles for ambulatory treatment. Jpn J Rhinol 49: 138–142, 2010 (in Japanese)
- Sygula-Cholewińska J, Lech T, Szostak-Kot J et al: ATP bioluminescence method in surface hygiene monitoring. https://www.researchgate.net/publication/270283798\_A TP\_BIOLUMINESCENCE\_METHOD\_IN\_SURFACE \_HYGIENE\_MONITORING (accessed on Nov 16, 2017)
- 4) 3M Clean-Trace<sup>™</sup> Product Instructions. 2013, http://mul timedia.3m.com/mws/media/807474O/3m-clean-trace-su

rface-atp-uxl100-product-instruction.pdf (accessed on Nov 16, 2017)

- 5) Spaulding EH, Cundy KR, Turner FJ: Part IV. Medical and health-related applications 33 Chemical disinfection of medical and surgical materials, infection. *In* Disinfection, Sterilization and Preservation, 2<sup>nd</sup> ed. (Block SS ed), pp654–684, Lea & Febiger, Philadelphia (1977)
- Ikeda K, Sakai Y, Misawa N et al: Bacterial contamination of multiple-use atomizers commonly used in Japan. J Jpn Soc Infect Aerosol Otorhinolaryngol 26: 51– 53, 2008 (in Japanese)
- Aycicek H, Oguz U, Karci K et al: Comparison of results of ATP bioluminescence and traditional hygiene swabbing methods for the determination of surface cleanliness at a hospital kitchen. Int J Hyg Environ Health 209: 203– 206, 2006
- Turner DE, Daugherity EK, Altier C et al: Efficacy and limitations of an ATP-based monitoring system. J Am Assoc Lab Anim Sci 49: 190–195, 2010